Telephone: 91-532-227+1122, 1123 : 0532-2545341, 2545822 Fax

GRAM : MNNIT

Website : http:\\www.mnnit.ac.in Email

: spo@mnnit.ac.in, fip@mnnit.ac.in



elschyly ug: jkVh, il6 l6xdh l lHku bylgkkn

इलाहाबाद-211004 (भारत)

Motilal Nehru National Institute of Technology Allahabad

Allahabad-211004 (India)

An Institute of National Importance as Declared by NIT Act, GOI, 2007

INVITATION FOR SUPPLY OF AUTO TITRATOR SYSTEM WITH KARL FISCHER TITRATION (For Contracts value estimated to cost less than Rs.25lakhs)

1	Closing Date/Time for submission of Quotations	25/11/2014; 15:00 Hrs.	
2	Opening Date/Time of Quotations	25/11/2014; 15:30 Hrs.	
3	Quotations to be submitted at	The Office of Faculty-In-charge (Purchase)	
		M.N.N.I.T. Allahabad, Allahabad-211 004 (U.P.)	
4	Place of opening of quotations	The Office of Faculty-In-charge (Purchase)	
		M.N.N.I.T. Allahabad, Allahabad-211 004 (U.P.)	
5	Tender Fee	Rs. 200.00 (Non-refundable), in form of a Demand	
Draft drawn in favor of DIRECTO		Draft drawn in favor of DIRECTOR, MNNIT	
Allahabad.		Allahabad.	
6	Earnest Money Deposit (EMD)	Rs. 35,000/- (Rs. In words Thirty five thousand	
		only), in form of FDR /Bank guarantee, from any of	
		the Commercial Banks, drawn in favour of	
		"Director, MNNIT Allahabad" payable at	
		Allahabad, and valid for a period of 45 days	
		beyond the final bid validity period	

To.

Dear Sir

You are invited to submit your most competitive quotation for the following goods/equipments. The quotation should be sent directly to the undersigned under Sealed Cover marked "Quotation Reference No., Date", and "The Due Date:

SI. No.	Brief Description of the Goods	Quantity.	Place Of Delivery	Installation Requirement
1.	Auto Titrator System with Karl Fischer Titration	01 No.	M. N. N. I.T.	Yes
	Detailed Specification attached at 'Annexure-B'		Allahabad	

The sealed tenders must be submitted in the Office of the Faculty In-charge (Purchase), on or before due date & time as mentioned above in the office of the undersigned, in the presence of bidder's authorized representatives, who choose to attend and shall sign a register evidencing their attendance.

The tender/quotations NOT submitted in the prescribed format or Incomplete in any respect will be outrightly rejected.

Faculty in-Charge, Purchase MNNIT Allahabad

GENERAL TERMS & CONDITION (GCC)

1. Bid Price

- (a) The contract shall be for the full quantity as described above. Corrections, if any, shall be made by crossing out, initialing, dating and rewriting.
- (b) All duties, taxes and other levies payable by the contractor under the contract shall be included in the total price.
- (c) The rates quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
- (d) The prices should be quoted in Indian Rupees only for indigenous items. For imported items/Equipments, the prices should be quoted in foreign currency.
- (e) Each bidder shall submit only one quotation.

2. Validity of Quotation

Quotation shall remain valid for a period not less than 90 days after the deadline date specified for submission.

3. Evaluation of Quotations

- (a) The purchaser will evaluate and compare the quotations determined to be substantially responsive, i.e., which
 - (i) are properly signed; and
 - (ii) conform to the terms and conditions, and specifications.
 - (iii) will be with tender fees, and EMD required, if any.
- (b) The Quotations would be evaluated separately for each item.

4. Award of Contract

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

- Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
- The bidder whose bid is accepted will be notified for the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order if falls under Govt. purchase rules, i.e., GFR-2005

5. Formats and Signing of Bid

- Each & every pages of the submitted bid must carry the page numbers.
- The bidders are not allowed to make additional and alteration in the tender document, any additions and alteration in the tender document, any additions and alterations shall be at bidder's risk. Conditional/modified tender are liable to be rejected.
- The tender/quotations NOT submitted in the prescribed format or Incomplete in any respect will be outrightly rejected.
- If the bid submitted without compliance sheet details of equipments the bid will be outrightly rejected.
- The bid shall be typed or written in indelible ink and shall be signed by the Bidder/Tenderer or a person or persons duly authorized to bind the Bidder/Tenderer to the Contract. All pages of the bid, except for unamended printed literature, shall be initialed by the person or persons signing the bid.
- Any interlineations, erasure or overwriting shall be valid only if they are initiated by the persons or persons signing the bid.
- The Bidder shall furnish information as described in the Form of Bid on commissions or gratuities, if any, paid or
 to be paid to agents relating to this Bid, and to contract execution if the Bidder is awarded the contract.
- Payment condition: 100% Payment shall be made only after satisfactory delivery and installation of the goods.
 Advance payment is not permitted.
- 7. Normal Commercial warranty/ guarantee shall be applicable to the supplied goods.
- 8. Payment will be cleared through e-payment mode (Format of e-payment form given at Annexure-D)

INSTRUCTIONS/GUIDELINES TO BIDDERS (ITB)

- 1. Bidders who had supplied the Government departments, PSUs/NITs/IITs and reputed Institutions and Organizations should only apply.
- 2. The bidder should enclose the cliental list with contact address along with Phone/Fax for reference, for the quoted item.
- 3. Please quote the prices of F.O.R. Motilal Nehru National Institute of Technology Allahabad, Allahabad.
- 4. The Bidders are requested to submit their Quotation with EMD in the form of Demand Draft/ Bank Guarantee/ Short term Deposit of the cost of equipments/items in favor of "The Director, MNNIT, Allahabad" payable at Allahabad. Quotation without EMD will be summarily rejected. EMD's of all Bidders will be returned back after the purchase order /contract agreement is issued to successful bidder.
- 5. Successful Bidders must have to submit the performance security @ 10% of the purchase order or contract value and it will be retained up to the warranty period.
- 6. The bidder should enclose relevant documents wherever necessary to substantiate his eligibility.
- 7. The Institute is eligible for and can provide necessary documentation for Custom/Excise and Octroi exemption on demand, if permitted under rule.
- 8. Motilal Nehru National Institute of Technology Allahabad, Allahabad, reserves the right to reject, accept any or whole or part of any of the tender without assigning any reason thereof and no claim will be heard. In case of dispute, the decision of The Director, MNNIT, and Allahabad will be final and binding.
- 9. If the successful bidder is not able to supply the goods items within the delivery date specified in the purchase order, the purchase order will be automatically treated as cancelled after expiry of delivery date period. The vendor will have to submit an undertaking in this regard that this condition is acceptable to him. In case of extra ordinary circumstances the vendors must send a request for extension of validity of purchase orders, with proper justification prior to the expiry of validity date for consideration.
- 10. In case of sophisticated equipments, an on-site training needs to be provided with no extra cost.
- 11. The quotation must be accompanied by leaflets/operation manuals/books etc.
- 12. Wherever applicable the vendors must quote the branded /desired items.

CHECK LIST (ON THE LETTER HEAD OF THE BIDDER)

The Bid will be the compilation of following documents, along with required supporting documents. No document in support of minimum eligibility criteria will be accepted / entertained after opening of tender.

SI. No	Documents	
Cover letter by bidder (On the Letter Head of the Bidder)		
2.	Tender Fee	
3.	Earnest Money Deposit (EMD), if required	
4.	Cliental List for quoted item	
5. Annexure-A: Bid Proposal Sheet (On the Letter Head of the Bidder)		
6.	6. Annexure-B: Compliance of bidder with reference to various equipments (On the Letter Head of the Bidder)	
7. Annexure-C: PRICE BID (On the Letter Head of the Bidder)		
8.	Annexure-D: PROFORMA FOR DIRECT PAYMENT/TRANSFER TO BANK ACCOUNT BY MNNIT ALLAHABAD	

ANNEXURE-A

BID PROPOSAL SHEET (ON THE LETTER HEAD OF THE BIDDER)

To

Faculty In-charge (Purchase) MNNIT Allahabad, Allahabad

Subject: Supply and Installation of "Auto Titrator System with Karl Fischer Titration" at MNNIT Allahabad, Allahabad.

Dear Sir,

We, the undersigned Tenderers, having read and examined in detail the specifications as specified in this document in respect of Supply and Installation of **Auto Titrator System with Karl Fischer Titration** at MNNIT Allahabad, Allahabad do hereby propose to supply the required products and services.

Tender No.					
Tender Fe	es submitted	YES / NO	(Please strike off whatever is not applicable)		
Amount	Mode	Date of Issue	Name of Bank	Valid up to	
	Demand Draft				
EMD submitted		YES / NO	(Please strike off what	ever is not applicable)	
Amount	Mode	Date of Issue	Name of Bank	Valid up to	
	Demand Draft				

- (i) ADDITIONAL PURCHASE/WORK ORDER: We understand that the MNNIT Allahabad, Allahabad, in case of the requirements may also place repeat purchase order/work order on the company. In such cases, we shall accept and execute all the purchase/work order placed on us by MNNIT Allahabad, Allahabad.
- (ii) **BID PRICING:** We further declare that the prices stated in our proposal are in accordance with your Terms & Conditions in the bidding document. We further understand that the quantities as specified in this Tender may increase or decrease at the time of Award of Purchase Order as per the requirements of MNNIT Allahabad, Allahabad.
- (iii) **QUALIFYING DATA:** We confirm that we satisfy the qualifying criteria and have attached the requisite documents as documentary proofs. In case you require any further information/documentary proof in this regard during evaluation of our bid, we agree to furnish the same in time to your satisfaction.
- (iv) CONTRACT PERFORMANCE SECURITY: We hereby declare that in case the contract is awarded to us, we shall submit the performance Guarantee Bond in the form of Bank Guarantee for the amount mentioned at clause No. 5 in ITB of the total order value.
- (v) PAYMENT TERMS: We hereby declare that in case the contract is awarded to us, we agree with payment terms specified in the tender documents.
- (vi) CERTIFICATE AND DECLARATION:
 - a) I/We certify that no addition/modification/alteration has been made in the Original Tender Document. If at any stage addition /modification /alteration is noticed in the Original Document, I/We will abide by the terms and conditions contained in the original tender document, failing which MNNIT Allahabad reserves the right to reject the tender and/or cancel the contract
 - b) It has been certified that all information provided in tender form is true and correct to the best of my knowledge and belief. We hereby declare that our proposal is made in good faith, without collusion or fraud. No forged /tampered document(s) are produced with tender form for gaining unlawful advantage. We understand that MNNIT Allahabad is authorized to make enquiry to establish the facts claimed and obtained confidential reports from clients.
 - c) In case it is established that any information provided by us is false / misleading or in the circumstances where it is found that we have made any wrong claims. Further MNNIT Allahabad is also authorized to blacklist our firm/company/agency and debar us in participating in any tender/bid in future.
 - d) I / We assure the Institute that neither I /We, nor any of my /our workers, will do any act which is improper / illegal during the execution in case the tender is awarded to us.
 - e) I / We assure the Institute that I / We will NOT be outsourcing any work specified in the tender document, to any other firm.
 - f) Neither I / We, nor anybody on my /our behalf will indulge in any corrupt activities /practices in my /our dealing with the Institute.
 - g) Our Firm / Company / Agency is not been blacklisted or banned by any Govt. Department, PSU, University, Autonomous Institute or any other Govt. Organization.
 - h) I/We certify that, I have understood all the terms & conditions (GCC and ITB etc.), as indicated in enquiry of the tender document, and hereby accept all the same completely.
 - i) I/We, further certify that I/We, possess all the statutory /non-statutory registrations, permissions, approvals, etc., from the Competent Authority for providing the requisite services,
 - j) We understand that you are not bound to accept the lowest or any bid you may receive.
 - k) I/We hereby declare that this tender on acceptance communicated by you shall constitute a valid and binding contract between us.
 - I) I/We certify that the submitted quotation is duly paginated and contains from page no. 1 to

Date: Signature and Seal of the Manufacturer/Bidder

Technical Compliance of Goods (ON THE LETTER HEAD OF THE BIDDER)

Potentiometric Auto Titrator System with Parallel Titration	SI. No.	Specification	Quantity	Compliance (Yes/No)	Quoted Model
Titration 1. The system should be capable of doing all types of Titrations like Acid base, Non Aqueous, Redox etc. 2. The system should also have the facility to perform Karl Fischer Titration. 3. The system should be completely software controlled. Capable of performing at least 2 titrations in parallel. 4. Possibility to connect up to Four Titrating burettes for different types of titrations should be available. 5. System should be capable of performing parallel titration as and when required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titratis without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWD galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable extra Magnetic/Propellar with control on stirring speed and direction should be provided. 15. Suitable externed swithout the cell having to be opened. 16. Latest windows based software should be provided with the system, with the following facilities. 18. It have been suitable to the provided and provided along with the instrument. 19. Latest windows based software should be provided with the system, with the following facilities. 20. It should allow to use the IF statements to do actions based on results. 31. Respect to the pr	1.	· · · · · · · · · · · · · · · · · · ·	01 No.	,	
1. The system should be capable of doing all types of Titrations like Acid base, Non Aqueous, Redox etc. 2. The system should also have the facility to perform Karl Fischer Titration. 3. The system should be completely software controlled. Capable of performing at least 2 titrations in parallel. 4. Possibility to connect up to Four Titrating burettes for different types of titrations should be available. 5. System should be capable of performing parallel titration as and when required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange I titrans without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 12, 5 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments Should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWG galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the tollowing facilities. 2. It should allov to use the IF statements to do actions based on results. 3.		· · · · · · · · · · · · · · · · · · ·			
Non Aqueous, Redox etc. 2. The system should also have the facility to perform Karl Fischer Titration. 3. The system should be completely software controlled. Capable of performing at least 2 titrations in parallel. 4. Possibility to connect up to Four Titrating burettes for different types of titrations should be available. 5. System should be capable of performing parallel titration as and when required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange I titrats without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to 17W0 galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. 15. Latest windows based software should be provided with the system, with the following facilities. 16. Latest windows based software should be provided with the system, with the following facilities. 2. It should allow to use the IF statements to do actions based on results. 3. Respect to reasonable to do two titrations parallely in the same window. 2. Result monitoring, customized report and automatic data back-up should be available.		Titration			
 The system should also have the facility to perform Karl Fischer Titration. The system should be completely software controlled. Capable of performing at least 2 flations in parallel. Possibility to connect up to Four Titrating burettes for different types of litrations should be available. System should be capable of performing parallel titration as and when required. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titrants without contamination. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent litration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-u					
3. The system should be completely software controlled. Capable of performing at least 2 livations in parallel. 4. Possibility to connect up to Four Titrating burettes for different types of titrations should be available. 5. System should be capable of performing parallel titration as and when required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titrants without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. 18. It should allow to use the IF stalements to do actions based on results. 19. Result monitoring, customized report and automatic data back-up should be available. 20. Reagent consumption, monitor, a warring should be displayed when the selected limit is reached.					
at least 2 litrations in parallel. 4. Possibility to connect up to Four Titrating burettes for different types of titrations should be available. 5. System should be capable of performing parallel titration as and when required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titrants without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent litration we seel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. 2. It should allow to use the IF statements to do actions based on results. 3. Result monitoring, customized report and automatic data back-up should be available. 2. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. 3. Audomatic inquiry of sample size and/or ID number after sta		3 ,			
titrations should be capable of performing parallel litration as and when required. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange "Itrants without contamination. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available.		at least 2 titrations in parallel.			
required. 6. The equipment should not have a separate method for drift determination in KF titration and the drift compensation should be automatic. 7. Burrette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titrants without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burrettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. 2. It should allow to use the IF statements to do actions based on results. 3. Result monitoring, customized report and automatic data back-up should be available. 2. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. 3. Automatic inquiry of sample size and/or ID number after start of titration on request. 4. Variable Start Delay for poorty soluble samples. 5. Should be possible to do two titrations parallely in the same window. 5. Result trend analysis, History about the each determination should be available		titrations should be available.			
KF litration and the drift compensation should be automatic. 7. Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange T itrants without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 mt should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorty soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.					
 Burette should have a minimum resolution of 1/10000 of its burette Volume and the flexibility to interchange Titrants without contamination. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warring should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Res		·			
and the flexibility to interchange Titrants without contamination. 8. For ease of handling automatic cleaning and presentation of burette with a single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		·			
 For ease of handling automatic cleaning and presentation of burette with a single command should be possible. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent litration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Reayent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 					
single command should be possible. 9. Burettes volume of 2, 5, 10, 20 and 50 ml should be available. 10. Dosing drivers with chip on board technology with the capability of read write. Should be provided. 11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. 12. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent litration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		· · · · · · · · · · · · · · · · · · ·			
 Burettes volume of 2, 5, 10, 20 and 50 ml should be available. Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 		· · · · · · · · · · · · · · · · · · ·			
 Dosing drivers with chip on board technology with the capability of read write. Should be provided. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 		· · · · · · · · · · · · · · · · · · ·			
 The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and performance of sensors. Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 		10. Dosing drivers with chip on board technology with the capability of read			
 Maximum up to TWO galvanically separated measuring interfaces should be provided for parallel operation of the instrument. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warring should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 		11. The instruments should have the facility to connect intelligent sensors with chip on board technology to capture all data related to calibration and			
provided for parallel operation of the instrument. 13. A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. 14. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration wessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		·			
 A suitable stirrer Magnetic/Propellar with control on stirring speed and direction should be provided. A suitable KF Titration Stand not only for stirring, but also for manually exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. Suitable electrodes and accessories required for the applications should be provided along with the instruments. Latest windows based software should be provided with the system, with the following facilities. It should allow to use the IF statements to do actions based on results. Result monitoring, customized report and automatic data back-up should be available. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. Automatic inquiry of sample size and/or ID number after start of titration on request. Variable Start Delay for poorly soluble samples. Should be possible to do two titrations parallely in the same window. Result trend analysis, History about the each determination should be available. 					
exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be aspirated or added without the cell having to be opened. 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		13. A suitable stirrer Magnetic/Propellar with control on stirring speed and			
 15. Suitable electrodes and accessories required for the applications should be provided along with the instruments. 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available. 		exchanging the spent titration vessel contents should be given along with the instrument. With the integrated membrane pump, solvent should be			
 16. Latest windows based software should be provided with the system, with the following facilities. a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available. 		15. Suitable electrodes and accessories required for the applications should be			
 a. It should allow to use the IF statements to do actions based on results. b. Result monitoring, customized report and automatic data back-up should be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available. 		16. Latest windows based software should be provided with the system, with the			
be available. c. Reagent consumption, monitor, a warning should be displayed when the selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		-			
selected limit is reached. d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		·			
d. Automatic inquiry of sample size and/or ID number after start of titration on request. e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		c. Reagent consumption, monitor, a warning should be displayed when the			
e. Variable Start Delay for poorly soluble samples. f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		d. Automatic inquiry of sample size and/or ID number after start of titration			
f. Should be possible to do two titrations parallely in the same window. g. Result trend analysis, History about the each determination should be available.		•			
available.		f. Should be possible to do two titrations parallely in the same window.			
17. Three years of extended warranty should be quoted.					
, , , , , , , , , , , , , , , , , , , ,		17. Three years of extended warranty should be quoted.			

PRICE BID (ON THE LETTER HEAD OF THE BIDDER)

SI.		Quantity	Quoted Unit Rate in Rs	Total Amount (in Rs.)	
No.				In Figures	In Words
	Auto Titrator System with Karl Fischer Titration				
	Detailed Specification attached at 'Annexure-B'				

We	e agree to supply the above goods in accordance with the technical specifications for a total contract price
of_	within the period specified in the Invitation for Quotations.
We	e also confirm that the normal commercial warrantee/guarantee of months shall apply to the offered goods.
	Signature and Seal of the Manufacturer/Bidder
Nο	te:
	Preference will be given to ISO 9001-2000 certified manufacture/supplier, who can ensure the manufacturing of the
	machine as per the required testing standards/tender specifications within the specified tolerance limit.
2.	Bidders should provide copies of original Memorandum and Articles of Association, defining the constitution of legal
	status, place of registration and place of business of the company.
===	

ANNEXURE-D

PROFORMA FOR DIRECT PAYMENT/TRANSFER TO BANK ACCOUNT BY MNNIT ALLAHABAD

SI. No.	Particulars	Information
1.	Firm [Beneficiary] Name	
2.	Code No. allotted by the MNNIT Allahabad	
	[If code is not allotted yet, please enclose a cancelled cheque and copy of PAN	
	card. Cancelled cheque & PAN card is to be submitted only once]	
3.	Complete Bank Account No. of the Firm [beneficiary].	
	[in case of change in bank account vendor should write to Account Office]	
4.	Bank Name	
5.	Branch Address	
6.	IFSC Code No.	
7.	Permanent Account Number	
8.	Mobile No. [for SMS]	
9.	Email ID [for information]	

We undertake that all information provided above is correct and MNNIT Allahabad will not be responsible in case of any error on the part of firm.

[Seal and Signature of the firm]